

EC CERTIFICATE
Full Quality Assurance System
Directive 93/42/EEC on Medical Devices, Annex II excluding (4)

No. 5-796-200-1609

National Institute of Pharmacy and Nutrition
Directorate of Device Testing and Clinical Engineering (EMKI)

certifies that the manufacturer:

INMED-Karczewscy sp. z o.o. sp. k.
ul. Kwiatowa 32A
55-330 Krępiec
Poland

for the products / product category:

Devices and the equipment for medical gases

applies a quality system which meets the requirements of Directive 93/42/EEC concerning medical devices, Annex II.

Registry number of the related audit report: **42-104-2007**

This certificate is valid until **2021-09-01** supposed that the results of the regular yearly surveillance audits are satisfactory.


Issued by EMKI as a Notified Body with identification number **1011**.

This certificate is valid only with the attachment.

Budapest, 2016-09-02


Deputy General Director




Head of EMKI



EMKI 1528

The authenticity and validity of the certificate are verifiable at EMKI.

ATTACHMENT TO EC CERTIFICATE

Page 1 of 1

Additional information for Certificate No. 5-796-200-1609

The certificate is valid for the following products:

Medical supply units

Terminal units and probes for compressed medical gases, vacuum and scavenging systems

Pressure monitoring boxes with area shut-off valve and alarm systems


Pressure flexible connections for use with medical gas systems


The detailed product list is kept by EMKI under No. 42-104-2007.

Issue: 1

Date: 2016-09-02

First issued: 2016-09-02


Deputy General Director


Head of EMKI



EMKI